

QuickSplit 610 Series

Adjustable Flow Splitter

Methods for Setting Low Split Flow



Figure 1: Low flow out



Figure 2: Inlet port



Figure 3: High flow out



Figure 4: Split flow Vernier scale on splitter body

Adjusting the Split Ratio

Split ratios are decreased by turning the adjustment knob clockwise and increased by turning the adjustment knob counterclockwise. The split ratio range can be changed by exchanging the replaceable resistor cartridge. The model 610 series adjustable splitter can be configured for either post column or pre-column splitting.

Setting the Split Ratio

Listed below are several options for setting the split ratio of an adjustable splitter. Choose the method that best suits your purposes.

Routine split ratio setting

Each flow splitter is shipped with a Manufacturing Test Log unique to that splitter. If you require a replacement copy, contact ASI and provide the splitter part number and serial number. Refer to Figure 5 for a sample copy.

Set the inlet flow and the Vernier scale setting to the value listed on Manufacturing Test Log. The scale is on the splitter body (Figure 4).

Note: Once it's set, the split ratio will remain constant regardless of changes to inlet flow rate or solvent viscosity. Regarding gradients with different viscosity mobile phases, the low split flow will remain constant for Post-column applications. Accuracy is $\pm 10\%$ of the listed value, on account of system-to-system variability.

Figure 5

Manufacturing Test Log for 610-PO10-04 (Sample)

S/N:xxxx

Customer : xxxxxx

Ship date : xx.xx.xxxx

Leak check at inlet at 4,000 psi

pass

Leak check at outlet at 4,000 psi

pass

Solvent = Water

Input flow rate = 1.00 mL/min.

Split Ratio Range = 5:1 to 100:1

Vernier scale setting	Capillary flow, mL/min	Split ratio	Back Pressure, PSI
19	0.0075	132.333	31
22	0.015	65.6667	69
25	0.0256	38.0625	122
28	0.0361	26.7008	178
31	0.0477	19.9644	243
34	0.0565	16.6991	286
37	0.0709	13.1044	357
39	0.0841	10.8906	423
41	0.1012	8.88142	509
43	0.1168	7.56164	589
45	0.1347	6.4239	681
46	0.143	5.99301	723
47	0.156	5.41026	791
48	0.1618	5.18047	824
49	0.1725	4.7971	884
50	0.1799	4.55864	920

Resistor cartridge value = xxxx psi/(mL/min)

Comments :

Note: The above data is for zero back pressure downstream from flow splitter. Back pressure generated by splitter will vary depending on solvent viscosity, however split ratio will not be affected.

Adjust Back Pressure

Use pressure drop across the flow splitter and Ohms Law.

$$P = L \times Q^2 \times V$$

Q2: Desired split flow, mL/min.*P*: Pressure drop across splitter, PSI*L*: Fluid resistor value PSI/mL/min. (from calibration sheet)*V*: Viscosity in centipoise

Setting the Split Ratio on Pre-column Applications

It is more useful to work with actual flow rate rather than split ratio. Basic procedure is as follows:

1. Measure the pressure drop resulting from column at desired flow rate, using the same mobile phase as used in the method.
2. Refer to the Manufacturing Test Log to find the correct setting to achieve the pressure drop found in step 1. See a sample test log, Figure 5. Set the vernier setting to this setting.
3. Connect the low flow side to the column, and the inlet to the pump.
4. Adjust the pump flow rate until the pressure found in step one is achieved. The flow through the column will now match the desired flow rate in step one.

The flow through the column can be adjusted either by changing the splitter setting, or by changing the pump flow rate. Note that during a gradient with different viscosity mobile phase, the flow rate will change as the viscosity changes - but this change will be repeatable.

Configuring for Custom Split Ratio Ranges

A variety of resistor cartridges can be ordered from ASI to configure the Adjustable Flow Splitter for split ratios other than standard. Please contact technical service at 800-344-4340 for assistance in selecting the correct resistor cartridges. We will gladly assist you in determining the best splitter configuration for your application.



www.hplc-asi.com

Phone Number: 800-344-4340/+1(country code) 510-669-2250

Rev. 08_14_2026